	Application No.	Applicant(s)
Notice of Allewshills	09/932,889	WERNZ ET AL.
Notice of Allowability	Examiner	Art Unit
	Yelena G. Gakh, Ph.D.	1743
The MAILING DATE of this communication appear All claims being allowable, PROSECUTION ON THE MERITS IS (herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGOR OF THE OFFICE	ars on the cover sheet with the co OR REMAINS) CLOSED in this app or other appropriate communication GHTS. This application is subject to	prrespondence address
1. This communication is responsive to the Amendment filed (08/30/04 and Interview of 11/08/04.	
2. The allowed claim(s) is/are 29,31 and 32.		
3. The drawings filed on 20 August 2001 are accepted by the Examiner.		
4. Acknowledgment is made of a claim for foreign priority unda All b)	been received. been received in Application No uments have been received in this note that communication to file a reply content of this application. ed. Note the attached EXAMINER'S reason(s) why the oath or declaration be submitted. In's Patent Drawing Review (PTO-94) Amendment / Comment or in the Offer Should be written on the declaration.	omplying with the requirements AMENDMENT or NOTICE OF on is deficient.
each sheet. Replacement sheet(s) should be labeled as such in the 7. DEPOSIT OF and/or INFORMATION about the deposit attached Examiner's comment regarding REQUIREMENT FO	of BIOLOGICAL MATERIAL mu	and he cultimitted at a co
Attachment(s)		
1. ☐ Notice of References Cited (PTO-892)	Notice of Informal Pate	
2. Notice of Draftperson's Patent Drawing Review (PTO-948)	6. Interview Summary (P	TO-413),
B. Information Disclosure Statements (PTO-1449 or PTO/SB/08), Paper No./Mail Date	Paper No./Mail Date _ 7. ⊠ Examiner's Amendme	<u>11/08/04</u> . nt/Comment
l. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	8. ⊠ Examiner's Statement 9. □ Other	of Reasons for Allowance

Application/Control Number: 09/932,889

Art Unit: 1743

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with George L. Snider on 11/08/04.

The application has been amended as follows:

Title (New): "Apparatus for treating cytological or histological specimens"

Claim 29. (Amended) An apparatus for treating cytological or histological specimens comprising:

a plurality of processing stations;

an object holder for carrying at least one specimen, wherein the object holder includes a pair of hooks on opposite sides of the object holder;

a transport device for delivering the object holder into and out of the processing stations, wherein the transport device includes:

a robot arm having a free end movable in three-dimensions;

a gripper mounted at the free end of the robot arm for releasably grasping the object holder, wherein the gripper includes a pair of slots for respectively receiving the pair of hooks of the object holder, each of the pair of slots having a downwardly-facing open end; an actuation member movable to and from a closure position wherein the actuation member closes the open end of each of the pair of slots; and a rotatable switching cross engaging the actuation member, wherein a rotational position of the switching cross determines whether the actuation member is in the closure position or away from the closure position, and wherein the switching cross is rotated by engagement thereof with the object holder;

wherein the gripper is actuable to grasp the object holder by lowering the free end of the robot arm to cause the gripper to engage the object holder while the object holder resides in one of the processing stations; and

wherein the gripper is actuable to release the object holder by lowering the arm

Application/Control Number: 09/932,889

Art Unit: 1743

end of the robot arm beyond a position at which the object holder resides in and is supported by one of the processing stations.

Cancel claim 30.

In Clam 31, change [30] to -- 29 --.

2. The following is an examiner's statement of reasons for allowance:

the prior art does not teach or fairly suggest an apparatus for treating cytological or histological specimens comprising a robot arm with a free end movable in three dimensions with a gripper mounted at the free end of this arm and having specific details recited in the amended claim 29.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Yelena G. Gakh, Ph.D. whose telephone number is (571) 272-1257. The examiner can normally be reached on 9:30 am - 6:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on (571) 272-1267. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Telu Hale

Yelena G. Gakh 11/8/04